



Hydrion® Buffer Certification Sheet
&
Certificate of Traceability and Conformity

pH Value 10.00

Mfg. Lot#	<u>222515</u>	Mfg. Date	<u>8-15-2015</u>
Exp. Date	<u>08-15-2020</u>	Tested By	<u>PARESH SHAH</u>

NIST/IUPAC Standards used for Calibration of pH Meters

Buffer #1 pH	<u>6.865</u>	NIST Standard Reference Materials Used:
		POTASSIUM DIHYDROGEN PHOSPHATE
		(NIST Ref MAT'L No. 186-I-d)
		DISODIUM HYDROGEN PHOSPHATE
		(NIST Ref MAT'L No. 186-II-d)

Buffer #2 pH	<u>9.18</u>	NIST Standard Reference Materials Used:
		SODIUM TETRABORATE DECAHYDRATE
		(NIST Ref MAT'L No.187c)
		(NIST Ref MAT'L No.)

pH Values Recorded:

pH Meter #1 10.00 pH Meter #2 10.01 pH Meter #3 10.02

Temperature of Buffer Solution Tested: 25° C

Buffer Powder Used For	Amount Used	Date Used	Sample Weight
<u> </u> Chemvelopes	<u> </u>	<u> </u>	<u> </u>
<u> X </u> Capsules	<u> </u>	<u> </u>	<u>1.00 GRAM.</u>

Comments: 150 KGS. TO ADH

This certificate is to certify that the above buffer salt here-in described has been analyzed with and/or compared to appropriate National Institute of Standards and Technology and IUPAC reference materials. The buffer salt tested is certified in accordance with MIL-STD-45662.

Certified By Paresh shah Signature  Date: 08-15-2015