



**Hydrion® Buffer Certification Sheet  
 &  
 Certificate of Traceability and Conformity**

pH Value 9.00

Mfg. Lot#	<u>228014</u>	Mfg. Date	<u>10-15-2014</u>
Date Tested	<u>10-15-2014</u>	Tested By	<u>PARESH SHAH</u>

**NIST/IUPAC Standards used for Calibration of pH Meters**

Buffer #1 pH	<u>6.865</u>	NIST Standard Reference Materials Used:
		<b>POTASSIUM DIHYDROGEN PHOSPHATE</b>
		(NIST Ref MAT'L No. 186-I-d )
		<b>DISODIUM HYDROGEN PHOSPHATE</b>
		(NIST Ref MAT'L No. 186-II-d )
Buffer #2 pH	<u>9.18</u>	NIST Standard Reference Materials Used:
		<b>SODIUM TETRABORATE DECAHYDRATE</b>
		(NIST Ref MAT'L No.187c )
		(NIST Ref MAT'L No. )

**pH Values Recorded:**

pH Meter #1 9.01      pH Meter #2 9.00      pH Meter #3 9.00

Temperature of Buffer Solution Tested: 25° C

Buffer Powder Used For	Amount Used	Date Used	Sample Weight
<input checked="" type="checkbox"/> Chemvelopes	<u>                    </u>	<u>                    </u>	<u>5.02 grams</u>
<input checked="" type="checkbox"/> Capsules	<u>                    </u>	<u>                    </u>	<u>1.01 grams</u>

Comments:

This certificate is to certify that the above buffer salt here-in described has been analyzed with and/or compared to appropriate National Institute of Standards and Technology and IUPAC reference materials. The buffer salt tested is certified in accordance with MIL-STD-45662.

Certified By Paresh shah      Signature       Date: 10-15-2014



**QP-FORM-1650-176**

**Version # 3 Prepared by SF Approved by PS 8/09/2013**  
**Buffer Certification Sheet and Certificate of Conformity**

---